

Work Order	3109.5
Setup-Code	180612-10290-2801-02



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and Effectiveness

Test Object:

Coated Leneta - Foil vs. Salmonella choleraesuis DSM11320

Page 1 of 7



Work Order	3109.5
Setup-Code	180612-10290-2801-02

Report on Findings

Client: Address:	Nano-Care Deutschland AG Alfred-Nobel-Straße 10 66793 Saarwellingen					
Work order no.:	3109.5					
Test object:	Coated Leneta Foil vs. Salmonella choleraesuis DSM11320					
Sample description:	Coated Leneta Foil					
Date of receipt of sample:	14.09.2018					
Type of test:	JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and Efficacy					
Test Germ:	Salmonella choleraesuis DSM11320					
Test laboratory:	QualityLabs BT GmbH					
Address:	Neumeyerstrasse 46a 90411 Nuremberg, Germany					
Setup-Code:	180612-10290-2801-02					
Sample material:	Leneta Foil					
No. of pages in report:	7					
Report on findings Place to the client: Recip	ient: Nano-Care Deutschland AG Harald Gerauer, Laboratory Director QualityLabs BT GmbH					
	Markus Zehe, Managing Director QualityLabs BT GmbH					
	EN 24.09.2014 TN Page 2 of 7					



Work Order	3109.5
Setup-Code	180612-10290-2801-02

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria $(1.25 \times 10^4 \text{ CFU} / \text{cm}^2)$ is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t₀ value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t₂₄ value).





Work Order	3109.5
Setup-Code	180612-10290-2801-02

Assessment of antimicrobial activity

A logarithmic germ reduction of \geq 3 log scales of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Page 4 of 7



Work Order	3109.5
	0.0000
Setup-Code	180612-10290-2801-02

References to Testconditions

Testconditions				
Sample size	25	cm ²		
Foil size	16	cm ²		
Volume Inoculum	400	μl		
Sample cleaning	Isopropanol	-		

References to deviations, preincubations, special test conditions

NONE

Page 5 of 7



Work Order	3109.5
Setup-Code	180612-10290-2801-02

Test Results

	Sample Name	Sample Code		t ₀ (cells/cm ²))	t	₂₄ (cells/cm ²))	Reduction [%]	Log Reduction
1	Leneta-Foil P121-10 (Reference)	102900806180001	6,6 x 10 ⁴	5,5 x 104	7,1 x 10 ⁴	7,8 x 10 ⁴	7,3 x 10 ⁴	1,2 x 10⁵	-	-
2	Liquid Guard clean + primer + wipe	102900806180002				< 1,0 x 10 ¹	< 1,0 x 10 ¹	< 1,0 x 10 ¹	> 99,99	> 4
3	Liquid Guard pro + wipe	102900806180003				< 1,0 x 10 ¹	< 1,0 x 10 ¹	< 1,0 x 10 ¹	> 99,99	> 4

*see "Interpretation of Results", page 6

Test strain	Salmonella choleraesuis DSM11320			
Initial cell count inoculum / cm ²	1.25 x 10⁴			
Initials of the editor	OS			
Measurement ended on	15.06.2018			



Setup-Code 180612-10290-2801-02	Work Order	3109.5
Setup-Code 180612-10290-2801-02		0.0000
	Setup-Code	180612-10290-2801-02

Comments on test objects

NONE

Interpretation of the results based on the measurements

NONE

Editor: Mr. Shendi _____

Crosschecked: Mr. Zehe____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and Efficacy

EN 24.09.2014 TN

Please note: The reproduction of any part of this test report is not permitted without the written permission of QualityLabs BT GmbH. The result apply only to the samples tested.

Page 7 of 7